



TWL1500 Mekanik Tork Kalibrasyon Tezgağı

- Ölçüm aralığı : 1 - 1500 Nm
- Uygun olduğu standartlar:
ISO 6789:2003
BS EN 26789:2003
GGG-W-686 (Amerika Askeri Standartı)
- FMT serisi tork sensörleri ile birlikte kullanılır
- 3 farklı FMT tork sensörü için yuva mevcuttur
- Torkmetrenin hareket ettirilmesi yöntemiyle ölçüm yapar
- Tork sensörleri \pm %0,5 hassasiyet ile ölçüm yapar
- TTT veya T-BOX tork kontrol ünitesi ile çalışır
- Türkiye'de servis, yedek parça ve kalibrasyon hizmeti



Set Torque		Min	Max	Actual Readings			
10.0		9.62	10.42	10.10	10.09	10.10	10.09
30.0		28.85	31.25	29.51	29.48	29.45	29.47
50.0		48.08	52.08	50.18	50.17	50.16	50.14

The limits shown and the test equipment used for this calibration comply with the requirements of: ISO 6789 - 2003

The uncertainty of the test equipment at 20°C is $\pm 1\% k = 2$.

The test equipment used in the performance of the above calibration has international traceability through the following calibration laboratory which is UKAS accredited to ISO 17025:2005

UKAS Laboratory No.: 0256

Tester Model: Static TD

Serial No.: 66892


Cert No.: 66892

Date of Calibration
8 Nov 2013

Quality Manager
J.Source

TWL1500 model kalibrasyon tezgahı ile birlikte kullanılabilir tork sensörleri

Norbar #	Tork Sensörü		Ölçüm Aralığı Nm	Torkmetre Boyu (mm)	
	Model	Parça No.		Min.	Max.
Pos 1	FMT 10	50672.LOG	0,5 - 10	145	1310
	FMT 25	50673.LOG	1,25 - 25	145	1310
Pos 2	FMT 150	50674.LOG	7,5 - 150	240	1405
	FMT 400	50675.LOG	20 - 400	240	1405
Pos 3	FMT 1500	50676.LOG	30 - 1500	336	1500



Pro-Test & TWL1500



FMT Serisi Tork Sensörleri





TTT Serisi Tork Kontrol Ünitesi



T-Box Tork Kontrol Ünitesi



Pro-Test Kalibrasyon Ekipmanı



Surkon Makina Sanayi ve Ticaret LTD. ŞTİ.

İvedik Organize Sanayi Bölgesi - Özpetek Sanayi Sitesi 1390 Sokak No: 12 - TR-06378 - Yenimahalle / ANKARA
Tel: +90 (312) 386 14 64 / Faks: +90 (312) 386 14 88 / info@surkon.com / www.surkon.com



ISO 9001: 2008
Quality Management System
TS EN ISO / IEC 17021
AB-0049-Y5
Certificate No: 57812

SURKON
INTERNATIONAL